

MASS SPECTROMETRY REPORT



Sample Description

Analyst: SHAO
Sample: M7526-Linaclotide-02
M.W.: 1526.8
Lot. No.:

Instrument

Waters 2695
Probe: ESI
Nebulizer Gas Flow: 1.5L/min
CDL: -20.0v
CDL Temp.: 250 °C
Block Temp.: 200 °C

Probe Bias:

+4.5kv

Detector:

1.5kv

T. Flow:

0.5ml/min

B. Conc.:

50%H2O/50%ACN